

CITO-SYSTEM GmbH

Haimendorfer Straße 37 + 46
90571 Schwaig bei Nürnberg
Deutschland

Test des Reinigers CITOCLYN Uni Eco der Firma CITO-SYSTEM GmbH

Sehr geehrte Damen und Herren,

Hiermit bestätigen wir, dass das Reinigungsmittel **CITOCLYN Uni Eco** der Firma CITO-SYSTEM GmbH in der uns übergebenen Form im März 2021 bei einem Langzeittest mit 36.000 Nutzungszyklen keine auffallenden Beschädigungen an den in unseren Anlagen verwendeten Stopp-Ventilen hervorgerufen hat. Dies entspricht unserem üblichen Testrahmen, der auch mit anderen Reinigern durchgeführt wurde. Somit kann das getestete Reinigungsmittel im Rahmen dieser Nutzungsdauer in unseren Anlagen **Final Plate Cleaner** und **Auto-Clean System** verwendet werden.

Test of the cleaner CITOCLYN Uni Eco from CITO-SYSTEM GmbH

Dear Sirs,

We hereby confirm that the cleaning agent **CITOCLYN Uni Eco** from CITO-SYSTEM GmbH in the form handed over to us in March 2021 did not cause any noticeable damage to the stop valves used in our equipment in a long-term test with 36,000 usage cycles. This corresponds to our standard test framework, which was also carried out with other cleaning agents. This means that the tested cleaning agent can be used in our systems **Final Plate Cleaner** and **Auto-Clean System** within the above mentioned usage cycle.

Wartenberg, 24.03.2021



Michael Reuel
Managing Director



Boris Kröll
Managing Director

DUO-TECHNIK GmbH
Im Tiegel 4
36367 Wartenberg